Advanced Nuclear Science and Technology Techniques Workshop



Contribution ID: 100

Type: Oral

Design and construction of a gamma-ray spectrometer with water shielding for low-level natural occurring radioactive material measurement

Tuesday, 17 May 2022 10:45 (20 minutes)

Gamma-ray spectrometer with a single HPGe or NaI:Tl detector shielded with lead is often used to measure the activity concentration of radionuclides in soil samples. A passive water shield to reduce background radiation reaching the detectors was designed using GEANT4 Monte Carlo simulations and then constructed. IAEA-375 soil and beach sand each placed in Marinelli beaker were measured for 48 hours using two LaBr₃:Ce detectors placed inside the constructed water shield. The samples were also measured for 24 hours using a NaI:Tl detector inside the constructed water shield and HPGe shielded with lead and copper to compare and validate the results from measurements with the LaBr₃:Ce detectors. Both the simulated and measured results show that the water shield attenuates the 2614.5 keV gamma rays by over 90 % and energies lower than the 2614.5 keV by far above 90 %. The activity concentration of ⁴⁰K radionuclide in IAEA-375 soil and beach sand measured activity of the detector. The measured activity concentrations of ²³⁸U and ²³²Th series and ⁴⁰K radionuclides in IAEA-375 soil were comparable with certified values to within measurement uncertainties. The activity concentrations of ²³⁸U and ²³²Th series and ⁴⁰K radionuclides in beach sand were determined using all the measurement geometries and consistent to within 1 σ to 2σ level.

Primary author: Dr BASHIR, Munirat (Ibrahim Badamasi Babangida University, Lapai. Nigeria)
Co-authors: Prof. NEWMAN, Richard (Stellenbosch University); Dr JONES, Pete (iThemba LABS)
Presenter: Dr BASHIR, Munirat (Ibrahim Badamasi Babangida University, Lapai. Nigeria)
Session Classification: Environmental Applications