Contribution ID: 21 Type: Oral

Sub-Millimetre Particle Tracking Using PEPT with the PolarisJ Semiconductor Array

Tuesday, 19 March 2019 15:15 (15 minutes)

The Positron Emission Particle Tracking (PEPT) technique has been in development in Cape Town since 2009, and allows one to track a 1mm positron-emitting point source travelling at 1m/s to within 1mm, 1000 times a second. Typically, this utilises a large, high efficiency scintillation detector like BGO. However, recent experiments have shown that high resolution, relatively high efficiency semiconductor (CdZnTe) detectors, which are usually used in prompt gamma-ray detection during proton radiotherapy, are able to track particles down to sub-millimetre precision. We have been using the PolarisJ detectors to develop the PEPT technique, and hope to use them to track particles in micro-scale fluid dynamics applications. Initial measurements have been able to locate a ²²Na button source in three-dimensional space with an uncertainty of 0.11mm and a signal-to-noise ratio of 85%. This is a promising first step towards tracking micrometre-sized particles with sub-millimetre accuracy.

Primary authors: HYSLOP, Nicholas; PETERSON, Steve (University of Cape Town); LEADBEATER, Thomas

(University of Cape Town)

Presenter: HYSLOP, Nicholas

Session Classification: Metrology & Applications

Track Classification: Metrology and Applications